



West Bengal Tenders		eProcurement System of Government of West Bengal Published Corrigendum Details			
		Date : 15-Jan-2020 08:26 PM			
		 Print			
Organisation Chain :	HEALTH AND FAMILY WELFARE DEPARTMENT CHIEF MEDICAL OFFICER OF HEALTH - ALIPURDUAR				
Tender ID :	2020_HFW_263808_1				
Tender Ref No :	DHFWS/APD/19-20/NO-1226				
Tender Title :	PROCUREMENT OF EQUIPMENT FOR ASHRAM HOSTEL ALIPURDUAR				
Corrigendum Type :	Other				
Corrigendum Document Details					
Corr.No.	Corrigendum Title	Corrigendum Description	Published Date	Document Name	Doc Size(in KB)
1	CORRIGENDUM	SPECIFICATION OF EQUIPMENTS	15-Jan-2020 08:24 PM	Corrigendum.pdf 	590.11



Government of West Bengal
Office of the Chief Medical Officer of Health Alipurduar
& District Health & Family Welfare Samiti, Alipurduar

Registration No: S/M/2347 of 2015-16

Matrisadan Building, 1st Floor, New Alipurduar, Ward No-XVI, Dist-Alipurduar, Pin: 736121

Tele: 03564-257200, email: cmohapd@gmail.com

Memo No. DHFWS/APD/19-20/No- 1304

Dated-15/01/2020

Corrigendum

All concerned are hereby informed through this corrigendum regarding detailed specification of procurement of equipments for 11 no of Ashram Hostel, Alipurduar vide NleT no-DHFWS/APD/19-20/No-1226 ,Dated-03/01/2020

SL- No	INSTRUMENTS NAME	Specification
1.	HAEMOGLOBINMETER DIGITAL (ISO CERTIFIED) HAEMOGLOBIN STRIPS (FOR DIGITAL METER)	Digital Hemoglobinometer Includes follows item: a) HB hemometer b) HB Lancing Device c) OHB user manual d) Warranty card e) Power source: It has two 3.6 volt lithium ion battery (not inbuilt) 1. 1. Display Screen: - Display your test result, date, time, etc. 2. On/Off/Select button:- to turn the meter on and off, and to select displayed items. 3. Test Strip Slot:- for inserting the strip 4. Display Indicators:- DD-MM: Shows date and month Batch Code: Indicates the code number of test strips. 5. Battery symbol:- Indicates the battery level (Desirable) 6. Strip Symbol: Indicates when to insert test strip. 7. Blood Drip symbol: This symbol appears when the meter is ready for you to apply a blood drop. 8. Test Result Area: Displays the hemoglobin level. 9. Measurement Units: indicates the units for the test result (g/dl).
2.	DIGITAL SYGMOMANOMETER	Measurement Method: Oscillometry Measurement Position: Upper Arm Measurement Range: Pressure:0~290mmHg(0~36.0kPa), Pulse Rate:40~240bpm, Measurement Accuracy: Blood Pressure: ± 3 mmHg, Pulse Rate: large of 3bpm and ± 3%, Power Supply 4 AA Size Alkaline Battery or optional AC/DC Adapter Configuration Standard Conf.: Adult cuff, user manual, 4 AA Size alkaline battery with inbuilt tube and Nibpcuff
3.	STETHOSCOPE (Adult)	1. High quality sound 2. Freq range > 20hz – 2000hz 3. Standard headphones (AV part of headset) 4. Good quality tubing
4.	WEIGHT MACHINE (Digital)	Weighing Scale (Adult-125 KG), digital display good quality with 4AA size Alternative battery

5.	DIGITAL THERMOMETER (for clinical use)	Measurement Range: 32°C (89.6°F) ----- 42°C (89.6°F), Low Temperature Display: Temperature < 32°C (89.6°F) Display L°C(L°F), High Temperature Display: Temperature > 43°C (109.4°F) Display H°C(H°F), Accuracy: 35°C (95°F) ----- 39°C (102.2°F) ±0.1°C (0.2°F), Normal Beep Sound: 10 long beep sounds in 10 seconds, Fever Beep Sound: (>37.8°C/100°F): 30 short beep sounds in 10 seconds, Auto Power Off: 8 minute ±2 minute, Battery: AG3 or LR41, Life Of Battery: 2 years or 2500 times, Dimension: 125mm x 19mm x 11mm, Unit
6.	STADIOMETER	a) Height Stand Disassembles Into 4 Sections & Easy To Transport In Disassembled Condition. Suitable For Height Measuring At Any Location, b) Accurate, Portable, Durable And Light Weight With All Plastic Parts, c) Height Measuring Scale, Floor Model Without Weighing Scale, d) Measuring Range : 20 - 210Cm With 1Mm Graduation. Measuring In Inches And Centimetres
7.	SNELLEN CHART	a) None reflecting, matte finish eye Charts, Designed for Vision testing at 20 Feet distance, b) Eye Chart measures 24 x 11 in.
8.	Near vision chart (EYE CHART)	Near Vision Chart: a) Notations for vision testing at 40 cm (16 in.) b) Visual Acuity range 8M to .20M (20/200, 6/120 to 20/10, 6/3 equiv.) c) 40 cm (16 in.) cord to maintain proper viewing distance d) Eye Chart Measures 18 X 23 cm (7 X 9 in.) e) Plastic Washable Sheet, f) Wall Hanging Whole for easy hanging.
9.	ISHIHARA CHART	a) 38 Plate test Book, b) Accepted by leading authorities as a simple and accurate method for discovering congenital color blindness and red-green blindness, each in two forms, complete and incomplete, c) Includes four special plates for tests to determine the kind and degree of defect in color vision, d) Makes use of the peculiarity that in red-green blindness, blue and yellow appear remarkably bright compared with red and green, e) Color plates encased in specially designed album-type books for ease of handling. Detailed instructions included.
10.	BMI CHART	a) Standard quality, ISO certified, A3 Size, b) Plastic Washable Sheet, c) Wall Hanging Whole for easy hanging

C.M.O.H. & Secretary
D.H.&F.W.Samiti, Alipurduar.

Dated-15/01/2020

NIEt No. DHFWS/APD/19-20/No- 1304/1(8)

Copy forwarded for information and necessary action for wide circulation:

1. The Hon'ble Sabhadipati, Alipurduar Zilla Parishad, Alipurduar
2. The District Magistrate, Alipurduar
3. The Addl District Magistrate, Alipurduar (Health)
4. The Dy. Chief Medical Officer of Health, I/II/DMCHO/DPHNO, Alipurduar
5. The Accounts Officer, Office of the CMOH, Alipurduar.
6. The District Informatics Officer, NIC, with a request to upload the e-tender notice in the District Website
7. The System Coordinator, IT Cell, Dept. of H.F. Welfare, with a request to upload the e-tender notice in the State Health Website
8. Office copy

C.M.O.H. & Secretary
D.H.&F.W.Samiti, Alipurduar.